

- ⑥ Disconnect used drainage bag and immediately connect new sterile bag to the catheter or sheath
- ⑦ Avoid touching the ends of any connecting area
- ⑧ Use the new protective cap to cover the end of the used drainage bag
- ⑨ Empty the used drainage bag and dispose of according to local policies and guidelines
- ⑩ Remove gloves and discard as normal clinical waste
- ⑪ Always use an aseptic and non-touch technique


Prosys® Non Latex Gloves

- Open Prosys® pack to reveal 1 pair of sterile gloves
- Take care not to touch the inner contents
- Decontaminate hands before and after procedure by washing and drying thoroughly
- Follow the principles of an aseptic and non-touch technique
- Avoid touching the ends of any connecting areas
- Always adhere to local policies and guidelines



If the package is damaged, the device must not be used. This device is not suitable for re-sterilisation.

IMPORTANT INFORMATION

 **Reuse Precaution:** This item is a single use device. Do not re-sterilise, reuse or repackaging may create a risk of patient or user infection.



Contains or presence of Phthalates: This product contains the phthalate DEHP. Phthalates are widely used as plasticizers in medical products and have undergone extensive testing. However when using the device, it should be taken into consideration that a lack of evidence regarding DEHP and any disease or adverse effect does not mean that there are no risks.



CLINISUPPLIES
Ideas into solutions



CliniSupplies Ltd, Qualitas House, 100 Elmgrove Road, Harrow, Middlesex HA1 2RW, UK
Telephone: 020 8863 4168 Fax: 020 8426 0768 Email: info@clinisupplies.co.uk
www.clinisupplies.co.uk

Prosys is a trademark of CliniSupplies Ltd.

July 2013



Instructions for use

- ① Open Prosys® sterile pack taking care not to touch the inner contents
- ② The sterile gloves will now be on top of the drainage bag
- ③ Wash and dry hands thoroughly
- ④ Place on and wear the sterile gloves
- ⑤ Position and prepare the new Prosys® sterile bag do not remove the protective cap at this stage